



Running Contract Details	
Equipment Name	Nebulizer-Piston Compression
Running Contract Valid Till	03-06-2027
Tender Ref No	KMSCL/EP/T553/82/2024
Tendered Quantity	100
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Nebulizer-Piston Compression</b> <i>Model &amp; Make : Lifeline Alpha / Lifeline Medical</i>	2060.06 Incl.GST :12%	151.98	2212.04
		<b>2060.06</b>	<b>151.98</b>	<b>2212.04</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Nebulizer-Piston Compression**

#### Technical Specification

#### **Equipment: Nebulizer**

1. Should be Piston Compression Type
2. Should be light weight, portable and compact.
3. Should have a dust filter.
4. Should be able to deliver a flow rate ? 7 lpm
5. Should have air pressure ? 35 psi.
6. Should have a check valve to protect the device against contamination due to backward inhalation
7. Should be compatible for continuous use
8. Should works on 200-240Vac/50Hz.
9. Should be supplied with nebulization accessory kit with mask for adult and paediatric – 2 nos. each
10. Nebulization mask for adult and pediatric – 10 nos. each
11. Should be heavy duty type Nebulizer suitable for usage in Govt. Hospitals