

Running Contract Details	
Equipment Name	Dialysis Chair
Running Contract Valid Till	27-09-2026
Tender Ref No	KMSCL/EP/T528/58E/2024
Tendered Quantity	500
Supplier Name	M/s IMEDFURNS PRIVATE LIMITED
GST No	33AAGCI1214D1ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
61/1A SERVICE ROAD PERIYAPANICHERRI CHENNAI -600128	Contact Person	MR.PRADEEP KUMAR
	Phone	
	Mobile No	9500120662
	Email	tenders@imedfurns.com ,info@imedfurns.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Dialysis Chair Model & Make : IMED4032K/IMEDFURNS	71980 Incl.GST :18%	5038.6	77018.6
		71980	5038.6	77018.6

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Dialysis Chair

Technical Specification

Equipment: Dialysis Chair

1. Should have motorised backrest, height adjustment and trendelenburg
2. Should have dimension 1870X970 mm approximately.
3. Should have minimum seat width 620 mm.
4. Should have minimum height 640mm.
5. Should have one touch key for CPR.
6. Should have Patient load capacity of minimum 130 Kg,
7. Pair of removable, adjustable armrest with back & forth sideways and up-down movements
8. Cardiac chair and emergency head down position available
9. Should have a minimum 125 mm diameter twin wheel high grade synthetic body castors with central locking mechanism
10. Should have rusting free body in dialysis surroundings
11. Wired handset control
12. Should have integrated food/ reading table
13. Should have IV pole
14. Should have Battery Backup