



Running Contract Details	
Equipment Name	Infant CPR Mannequin
Running Contract Valid Till	11-01-2028
Tender Ref No	KMSCL/EP/T573/844/2025
Tendered Quantity	100
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Infant CPR Mannequin <i>Model & Make : LITTLE BABY / LAERDAL MEDICA</i>	29382 Incl.GST :18%	2056.74	31438.74
		29382	2056.74	31438.74

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Infant CPR Mannequin

1. The manikin should be a low-cost, lifelike infant CPR trainer
2. Should have realistic chest compressions and chest rise that allow students to learn proper techniques
3. Must have oral and nasal passages that allow realistic nose pinch required for mouth-to-nose ventilation
4. Must have natural obstruction of the airway that allows students to learn the important technique of opening the airway
5. Must have head tilt/chin lift and jaw thrust that allows students to practice correctly all maneuvers necessary when resuscitating a real patient
6. Should have a realistic chest compliance that means students can experience the proper technique required for chest compressions on infants. Must have foreign-body airway obstruction feature that allows the release of a foreign-body obstruction to be practiced through back blows and chest-thrust techniques
7. Must be supplied with economical disposable airways for quick and easy clean up.
8. The manikin must be equipped with means of providing feedback on performed CPR &ventilation on a battery-operated electronic feedback unit.
9. Should be supplied with removable and reusable faces for convenient and affordable maintenance