



Running Contract Details	
Equipment Name	Point of care ABG machine
Running Contract Valid Till	27-02-2028
Tender Ref No	KMSCL/EP/T569/17A/2025
Tendered Quantity	100
Supplier Name	M/s Biocare Diagnostics
GST No	32AAOFB5438P1ZZ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
787a-1 First Floor Kamala Building Chalappuram (Po) Calicut 673002	Contact Person	Linesh.T
	Phone	
	Mobile No	9847009640
	Email	mail@biocarecalicut.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Point of care ABG machine <i>Model & Make : ISTAT 300 / ABBOTT</i>	268800 Incl.GST :5%	21145.6	289945.6
2	Catridge I	661.5 Incl.GST :5%	52.04	713.54
3	Cartridge II	429.45 Incl.GST :5%	33.78	463.23
		269890.95	21231.42	291122.37

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Point of care ABG machine							
Labour	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00
Comprehensi ve	24,000.00	24,000.00	24,000.00	24,000.00	24,000.00	24,000.00	24,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Point of care ABG machine

Portable single hand held system for blood gases, lactate, heamatology, chemistries, electrolytes, coagulation and cardiac markers

1. A single portable hand held system weight of the device less than 1 Kg capable of measuring blood gas, lactate and other parameter as mentioned below.
2. It should give the results of GLUCOSE, Na, K, CHLORIDE, IONISED CALCIUM, HCT, PO₂ and PCO₂, Lactate and Creatinine. These tests should be done using not more than 5 catridges. Each test should be provided with calibrators. Price of the catridge will be taken for evaluation.
3. It should have at least 250 test result storage facility.
4. It should have facility to take print outs.
5. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission /FDA(US)/STQC CB certificate /STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate /test report shall be produced along with the technical bid.
6. The system should be operated either through the rechargeable battery or 230Vac with battery backup. In case of rechargeable battery additional one set battery and charger to be provided
7. The catridge should have at least 3 months shelf life from the date of supply of hospitals. The bidder shall also replace the catridges nearing expiry at free of cost on intimation from hospital.