



Running Contract Details	
Equipment Name	Stretcher trolley higher end with CPR Facility
Running Contract Valid Till	30-06-2026
Tender Ref No	KMSCL/EP/T469/1219/2022
Tendered Quantity	15
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Stretcher trolley higher end with CPR Facility <i>Model & Make : PK-8000 / Paramount Bed India Pvt Ltd</i>	304440 Incl.GST :18%	21310.8	325750.8
		304440	21310.8	325750.8
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Stretcher trolley higher end with CPR Facility				
Labour	5,100.00	5,100.00	5,100.00	
Comprehensive	10,300.00	10,300.00	10,300.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Stretcher trolley higher end with CPR Facility

1. Should have 3 sectional mattress base with chromed or epoxy-coated steel frame.
2. Should have mattress base width of 700mm and mattress base length of 2000mm approximately ($\pm 5\%$ tolerance).
3. Should have an anti static mattress of size 650 mm x 1950 mm approximately.
4. Should have hydraulic height adjustment of 610 mm – 980 mm.
5. Should have anti static castors of size 200 mm with braking facility.
6. Should be able to X ray the patient from positions along the entire length and width of the trolley
7. Should have a 5th wheel for easy manoeuvrability.
8. Should have back section angle of $0^\circ - 70^\circ (\pm 5 \text{ degree tolerance})$
9. Should have leg section angle (gas-spring assisted) of $0^\circ - 45^\circ (\pm 5 \text{ o tolerance})$.
10. Should have trendelenburg of $14^\circ (\pm 2 \text{ o tolerance})$.
11. Should have reverse trendelenburg of $7^\circ (\pm 2 \text{ o tolerance})$.
12. Should have weight bearing capacity of at least 200 kg.
13. Should have place for fixing oxygen cylinder.
14. Should have detachable telescopic IV rods with four hooks and foldable side rails.
15. Body should be pre-treated and powder coated.
16. 1Should be supplied with standard accessories such as a. Patient restraining straps – 1 set (at least 2 Nos). b. Arm rest – 1 pair. c. Transfer board Type side rails – I pair