

Running Contract Details	
Equipment Name	Minor items for Sub Centers
Running Contract Valid Till	22-02-2026
Tender Ref No	KMSCL/EP/T479(R)/1618/2022
Tendered Quantity	6500
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Snellan Vision Chart <i>Model & Make : Paras Enterprises</i>	28 Incl.GST :12%	2.07	30.07
2	Near Vision Chart	44.8 Incl.GST :12%	3.3	48.1
3	Dressing forceps	288.96 Incl.GST :12%	21.31	310.27
4	Kidney Tray	268.8 Incl.GST :12%	19.82	288.62
5	SS Drum	1702.4 Incl.GST :12%	125.55	1827.95
6	Instrument Tray with lid	1691.2 Incl.GST :12%	124.73	1815.93
7	Artery Forceps	487.2 Incl.GST :12%	35.93	523.13

Item-wise Price Details				
8	Sterilizer small	4248 Incl.GST :18%	297.36	4545.36
9	SS basin	655.2 Incl.GST :12%	48.32	703.52
10	Ambu Bag ÃƒÂ¢Ã¢â€šÂ¬Ã¢â€šÂ¬ Pediatric	980 Incl.GST :12%	72.28	1052.28
		10394.56	750.67	11145.23

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Snellan Vision Chart

1. Snellan Vision Chart : vinyl coated, numbers, alphabets and e- chart – 1 No
2. Near Vision Chart : vinyl coated, numbers, alphabets and e -chart - 1 No
3. Dressing forceps - Thumb forceps,6 inch toothed or non toothed – 1 No, SS 410
4. Kidney Tray - Kidney Tray Medium 1. Stainless steel 304 grade , Size :10 inches – 1 No
5. SS Drum - SS Drum Medium 1. Stainless steel 304 grad, . Autoclavable up to 134 deg Celcius, Size : 9 x 12 inches – 1 No
6. Instrument Tray with lid - Instrument tray with lid - Medium Tray with Lid ,Stainless 304 grade with handle , Size: 12” x 10” x 2.5” – 1 No
7. Artery Forceps - curved 6 inch, SS 410 – 1 No
8. Sterilizer small : - 1 No
 - a. Should have 300mm length, 150mm breadth, 125mm height and 1KW heater coil.
 - b. Should be made up of seamless stainless steel sheet of grade 304.
 - c. Should have indicators and power cord.
 - d. Should operate on 220/230Vac, 50Hz single phase supply.
 - e. Should have ISI mark.
9. SS basin : 1 No
 1. Stainless steel 304 grade with lid
 2. Size : 15 inches
10. Ambu Bag – Pediatric- 1 No
 - a. Should have silicon rubber bellow to withstand autoclave at 134 deg. C
 - b. Should provide with autoclavable face mask & Oxygen connecting tube

- c. Should be supplied with a carry pouch
- d. It should have a bag volume of 500 ml. and a variation of ± 100 ml. will be accepted.
- e. Should have an expiratory resistance of 2.2cms of water
- f. Should have an inspiratory resistance of 3.3cms of water
- g. It should have controlled flow rates and ventilation, and with reduced airway pressure.
- h. Should have a port in the bag to connect oxygen with reservoir bag

Note: Bidder should quote all the items mentioned in the BOQ failing which the offer will not be consider for L1 evaluation.