



Running Contract Details	
Equipment Name	Trop T/I Analyzer
Running Contract Valid Till	03-12-2027
Tender Ref No	KMSCL/EP/T409/239/2021
Tendered Quantity	100
Supplier Name	M/s Roche Diagnostics India Pvt Ltd
GST No	33AAACA0266H1Z7
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
4th floor SKCL Harmony Square Door No.48 and 50 Prakasam street T.Nagar Chennai 600 017	Contact Person	Deepak Ranjan
	Phone	044-43900300
	Mobile No	919742746564
	Email	kalyana.chakravarthi@roche.com,sarath_babu.varayil@roche.com ,dipin.chako@roche.com ,sarath.nair@roche.com ,gokul.panicker@roche.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Trop T/I Analyzer <i>Model & Make : Roche Cobas j232/Roche Diagnostics</i>	112100 Incl.GST :18%	7847	119947
2	Test Cards	414.4 Incl.GST :12%	30.56	444.96
		112514.4	7877.56	120391.96

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Trop T/I Analyzer							
Labour	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00
Comprehensive	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Trop T/I Analyzer

1. Should be able to measure the high sensitive quantitative measurement of Troponin-T / Troponin I. Should be at least a third-generation assay.
2. It should be a portable unit and have an inbuilt battery backup.
3. It should have a parameter-wise QC checking facility.
4. Should have a memory capacity of a minimum of 500 test results with date time and comments.
5. Should have memory storage for results of minimum 200 cartridges.
6. Should have a minimum sample volume of approximately 300 microlitres.
7. Should have an operating temperature ranging from 18-30 degrees Celsius.
8. Should work on 200-240Vac 50Hz power supply.
9. It should have an internal printing facility or supply with an external thermal printer.
10. Should be supplied with 50 strips of Troponin T/I Parameter.
11. Should be provided with a calibration certificate issued by the manufacturer at the time of installation.
12. Should have a safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or a valid detailed electrical and functional safety test report from ERTL.