



Running Contract Details	
Equipment Name	Pacemaker (Temporary) - Dual Chamber
Running Contract Valid Till	19-03-2027
Tender Ref No	KMSCL/EP/T544/303B/2024
Tendered Quantity	10
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
43/2675A Dhanya Complex Sastha Temple Road Ernakulam North PO Kochi- 682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	Pacemaker (Temporary) - Dual Chamber <i>Model & Make : Pace 203/Osyпка Medical GmbH</i>			421075.2 Incl.GST :12%	31054.3	452129.5	
				421075.2	31054.3	452129.5	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Pacemaker (Temporary) - Dual Chamber							
Labour	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00
Comprehensive	30,000.00	31,500.00	33,075.00	34,730.00	36,465.00	38,300.00	40,200.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Pacemaker (Temporary) - Dual Chamber

Technical Specification

Equipment: Pacemaker (Temporary) Dual Chamber

1. Should be a Dual Chamber Pacemaker (Temporary) for bradycardia treatment before, during or after a surgery.
2. Stimulation burst and permanent stimulation should be available for high pacing rate.
3. Should be compact & easy-to-operate device, particularly suitable for emergency treatments.
4. Safety features, including automatic lead and battery check.
5. Should have continuous monitoring of the battery status.
6. Should have shock and water-resistant housing.
7. Should have back up pacing during battery change.
8. Should have Modes AOO, AAI, VOO, VII, DDD, DDI/equivalent, DVI
9. Should have pacing rate 30 – 200 ppm.
10. Should have fast pacing (Burst rate) of 80-200 ppm. Should have AV Delay 20 – 300 ms.
11. Should have Pulse Amplitude (A/V) – 0.1 – 17.0.
12. Should have Sensitivity (A/V) 0.4-10.0 mV (A) and 1.0-20.0 mV (V).
13. Should have minimum battery backup > 150 hours.
14. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
15. Temporary pacing wires- 2 nos should be supplied along with each unit