



| Running Contract Details | |
|--------------------------------|---------------------------|
| Equipment Name | LCD Vision monitor |
| Running Contract Valid Till | 07-12-2027 |
| Tender Ref No | KMSCL/EP/T574/645/2025 |
| Tendered Quantity | 100 |
| Supplier Name | M/s Stellar Surgi Systems |
| GST No | 32ADVFS0292G1ZU |
| Installation & Delivery Period | 8 Week(s) |
| Up-time / PM vist | 95% & 2 Visits per year |
| Warranty period | 2 Years |

| Supplier`s Details | | |
|--|-----------------|------------------------------|
| Address | Contact Details | |
| Door No:2/858 J E K Complex Eranhipalam P.O Calicut-673006 | Contact Person | Sajith Rahman |
| | Phone | |
| | Mobile No | 9895971500 |
| | Email | info@stellarsurgisystems.com |

| Item-wise Price Details | | | | |
|-------------------------|--|---|------------------------------------|----------------|
| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
| 1 | LCD Vision monitor <i>Model & Make : I Chart HD Smart / Appasamy</i> | 33600 Incl.GST :5% | 2643.2 | 36243.2 |
| | | 33600 | 2643.2 | 36243.2 |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) | | | | |
|--|----------------------|----------------------|----------------------|----------|
| Rate | 3 rd Year | 4 th Year | 5 th Year | |
| LCD Vision monitor | | | | |
| Labour | 1,270.00 | 1,323.00 | | 1,379.00 |
| Comprehen sive | 1,983.00 | 2,072.00 | | 2,165.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :LCD Vision monitor

Technical Specification

Equipment: LCD Vision Monitor

1. Should have LCD monitor of 18.5 inches which is remote operated
2. Should have the following charts: Snellen charts, Log Mar chart, Numeric, Alphabets, Dot chart, Contrast sensitivity, Pediatric chart, Masking test, ETDRS contrast Isihara colour test, Cartoon charts, Red Green or duro chrome, Peripheral acuity.
3. Should have an available distance of 6 feet, 8 feet, 12 feet, 15 feet and 20 feet
4. Should have eye Educational charts about Lasik, Cataract, Retina, Cornea, Contact lens
5. Should operate from 200 to 240v, 50/60Hz power supply
6. Should have certificate from a competent authority CE/FDA(US)/STQC C3 certificate/ STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid