

Running Contract Details	
Equipment Name	ENT OP Unit
Running Contract Valid Till	13-02-2027
Tender Ref No	KMSCL/EP/T473/485/2022
Tendered Quantity	25
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	ENT OP Unit <i>Model & Make : GEM SURGITECH</i>			386400 Incl.GST :12%	28497	414897	
				386400	28497	414897	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
ENT OP Unit							
Labour	3,500.00	3,500.00	3,500.00	3,500.00	3,500.00	3,500.00	3,500.00
Comprehensive	10,350.00	10,350.00	10,350.00	10,350.00	10,350.00	10,350.00	10,350.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :ENT OP Unit

1. It should have suction, sprayer, Mirror Rewarmer, Cold light source, Head light, Cable, Instrument Tray, Otoscope attached to camera (1.2 MP or above) with monitor (14 inch LED) and also having recording facility, Cautery Thermal
2. It should have Pneumatic Foot Operated Chair
3. Patient Chair Specification

Electrically operated, Electrical/ Hydraulic height adjustment, lift 20cm, with foot switch, Remote control. The upper part is easily all round swivelling and fixable by an arresting brake, the tall back rest is adjustable forward beyond the vertical line, backwards it can be declined slightly more than the horizontal line without any special assistance, the patient immediately sits down in an optimum position for examination by the synchronous coupling of the back rest to the foot support, the chair can be quickly changed into a long solid couch.

- | | |
|-----------------------|--|
| a. Width | - 60cm |
| b. Height | - (upper edge back rest, lowest position)120cm |
| c. Depth | - Total 75cm |
| d. Depth of the seat | - 42 cm |
| e. Height of the seat | - S2-72 cm |
| f. Power supply | - 220V/50Hz |

4. Side Cabinet: Fully air tight see through cabinet with X-ray view top, three fully openable draws it can be used as formalin movable cabinet with UV light Sterilization.