

Running Contract Details	
Equipment Name	PCR REAGENTS AND KIT
Running Contract Valid Till	21-01-2027
Tender Ref No	KMSCL/EP/T547/415AD/2024
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s GCC BIOTECH INDIA PRIVATE LIMITED
GST No	19AACCG9158JIZ3
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	0 Years

Supplier`s Details		
Address	Contact Details	
Joychandipur Patharberia Road Bakrahat South 24 Paraganas 743377	Contact Person	Dinesh B
	Phone	
	Mobile No	8431199156 8431199156
	Email	info@gccbiotech.co.in

Item-wise Price Details					
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Total Viral Nucleric Acid Extraction Kits (50 reaction kit)	260	1800 Incl.GST :0%	148.68	1948.68
			1800	148.68	1948.68

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly

to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

I. Covid-19 RT-PCR Kit

1. Real time PCR Kit for detection of SARS-Cov2 virus.
2. Should detect N gene, Orf gene and have internal control
3. Indicative pack size of 96 reactions per unit. The number of units may differ based on available pack size and quotes may be accepted accordingly
4. Should have minimum 1 year shelf life.
5. Should be open reagents compatible with routine Real Time PCR machines such as ABI/Biorad/Roche

2. Total Viral Nucleic Acid Extraction Kits (50 reaction kit)

1. Should be a viral nucleic acid extraction kit for extracting viral
2. DNA and RNA from blood and body fluids.
3. The kit should work with silica membrane spin column technology
4. Carrier RNA should be used in the kit to capture maximum amount of the Viral RNA from sample and should help viralRNA to escape from degradation by RNases
5. Yield of the Viral nucleic acid should be >90% recovery.
6. Should supply all required consumables like 1.5ml centrifuge tube for sample lysis, DNase/RNase free micro centrifuge tube for elute, separate collection tubes for all wash steps.

Preferable 50 reaction kits