

Running Contract Details	
Equipment Name	Diode Laser (Dermatology)
Running Contract Valid Till	29-08-2026
Tender Ref No	KMSCL/EP/T539(R)/1157/2024
Tendered Quantity	5
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	Diode Laser (Dermatology) <i>Model & Make : Capstone four Heads Diode</i>			1738700.32 Incl.GST :12%	128229.15	1866929.47	
				1738700.32	128229.15	1866929.47	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Diode Laser (Dermatology)							
Labour	46,800.00	46,800.00	46,800.00	46,800.00	46,800.00	46,800.00	46,800.00
Comprehensive	1,24,800.00	1,56,000.00	1,87,200.00	1,87,200.00	1,87,200.00	1,87,200.00	1,87,200.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Diode Laser (Dermatology)

Technical Specification

Equipment: Diode Laser (Dermatology)

1. Laser Wavelength: 808+755+1064nm
2. Fluence 10 - 80 j/cm². Pulse width 10-400ms.
3. Spot size 13.5mmX38.5mm, 13.5mmX28.5mm. 12.5mmX14.5mm and Round 8.5 mm.
4. Cooling -10C. Number of laser beams should be minimum of 10.
5. Frequency 1 – 10 Hz.
6. Hand power of up to 3000W.