

Running Contract Details	
Equipment Name	Bicarbonate Mixer
Running Contract Valid Till	18-12-2025
Tender Ref No	KMSCL/EP/T469/283/2022
Tendered Quantity	50
Supplier Name	M/s Meditech Corporations
GST No	32AWKPP1213A1Z9
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
DOOR NO.72/26-B2 NADEPALLY COMPLEX ELAMAKKARA PO-682026	Contact Person	Abdul Rahim
	Phone	
	Mobile No	7356829721
	Email	enquiry@meditechcorp.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Bicarbonate Mixer</b> <i>Model &amp; Make : DIA Mix / MEDITECH CORPORATIONS</i>	69900 Incl.GST :12%	5155.12	75055.12
		<b>69900</b>	<b>5155.12</b>	<b>75055.12</b>
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
Bicarbonate Mixer				
Labour	3,000.00	3,000.00	3,000.00	
Comprehensive	9,000.00	9,000.00	9,000.00	

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Bicarbonate Mixer**

1. Should have conical bottom type vessel of capacity atleast 100litres, made of Lipde (linear low-density polyethylene) material. Supporting stand should be made of Stainless steel – SS 304 grade.
2. Should have stainless steel rod grit, non wetted pedestal stirrer - Not applicable in case of pressure flow mixing technology.
3. Should have Marine style 3 blade propeller- Not applicable in case of pressure flow mixing technology.
4. Blade should be stainless steel – SS 316 L grade or Pressure flow mixing technology.
5. Should have at least 0.5 Hp, single phase electric motor, 230V AC. 50 Hz.
6. Motor Assembly shall be quickly removable for easy cleaning and autoclaving.
7. RO water inlet should be Cpvc with ball valve @ 105 ltrs.
8. RO water outlet should be Cpvc with ball valve @ 100 ltrs.
9. Solute should be Cpvc with ball valve @ 4 ltrs.
10. Drain should be Cpvc with ball valve @ 0 ltrs.
11. Should supply with necessary spares and accessories.
12. Should operate on mains 220-240Vac, 50 Hz single phase.