# KERALA MEDICAL SERVICES CORPORATION LTD

(A Government of Kerala Undertaking)
Thycaud P.O, Thiruvananthapuram - 14, Kerala.

Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647

Email:ep.kmscl@kerala.gov.in

CIN: U24233KL200TSGC021616, PAN: AADCK4029M, GSTIN: 32AADCK4029M1ZK

Running Contract Details				
Equipment Name	t Name Electrolyte Analyser			
Running Contract Valid Till	11-09-2026			
Tender Ref No	Ref No KMSCL/EP/T538/41/2024			
Tendered Quantity	100			
Supplier Name	M/s LabX			
GST No	32BPSPS6759C1ZE			
Installation & Delivery Period	8 Week(s)			
Up-time / PM vist	95% & 0 Visits per year			
Warranty period	2 Years			

Supplier`s Details						
Address	Contact Details					
Opp. SFS Tiara Apartments	Contact Person	Mr. Arun				
Near Burma Road Jn. Kumarapuram Medical College P.O	Phone	471 2447639				
Thiruvananthapuram	Mobile No	9446004066 ,9446002066				
	Email	info@labx.in ,sales@labx.in				

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Electrolyte Analyser  Model & Make : LabX/Qline Biotech Pvt Ltd	68204 Incl.GST :18%	4774.28	72978.28			
2	Reagent required to perform 1000 test/month	2670.61 Incl.GST :18%	186.94	2857.55			
3	Deproteinizer required to perform 1000 test/month	267 Incl.GST :0%	22.05	289.05			
4	Reagent	5443.2 Incl.GST :12%	401.44	5844.64			
5	Deproteiniser	996.8 Incl.GST :12%	73.51	1070.31			
6	Sodium Electrode	14750 Incl.GST :18%	1032.5	15782.5			
7	Potassium Electrode	14750 Incl.GST :18%	1032.5	15782.5			

Item-wise Price Details								
8	Refe	erence Electrode			15104 Incl.GST :18%	1057.28		16161.28
9	Tubi	ng Kit			3540 Incl.GST :18%		247.8	3787.8
10	Mem	brane Assembly			3304 Incl.GST :18%		231.28	3535.28
11	Chlo	oride Electrode			14750 Incl.GST :18%		1032.5	15782.5
					143779.61		10092.09	153871.7
Annual / Comprehensive Maintenance Charges (Exl.Tax)								
Rate	Rate 3 <sup>rd</sup> Year 4 <sup>th</sup> Year		4th Yea	ır	5 <sup>th</sup> Year			
Electrolyte Analyser								
Labou	r		7,800.00		7,	,800.00	7,800.00	
Compi	rehensi		28,800.00		28,800.00 28,800.0		28,800.00	

# **Other terms & conditions**

- 1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
- 2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
- 3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
- 4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
- 5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

# **Technical Specification**

# **Equipment : Electrolyte Analyser**

### **Equipment: Electrolyte Analyzer**

- 1. Should be able to measure sodium potassium, ionized calcium & chloride.
- 2. Should have a measuring method of Ion Selective Electrode (ISE).
- 3. Should be able to measure sodium and potassium in serum and body fluids.
- 4. Should have a throughput of minimum 60 samples per hour
- 5. Should have electrode for sodium, potassium, ionized calcium and chloride in case of electrode based technology
- 6. Resolution should be at least 0.1mmol/litre for each parameter

- 7. Should have automatic calibration, 1 and 2 point calibration, 2 point time bound (4<sup>th</sup>hour)calibration
- 8. Should have OC memory storage of at least 2 levels
- 9. Stand-by mode user controlled and automatic
- 10. Should have a measuring range for sodium 40 to 200 mmol/l, potassium 1.5 to 10 mmol/l
- 11. It should require 120 micro liter or lesser for whole blood serum
- 12. Should have printing facility
- 13. Should supply reagent pack for 1000 tests, one internal filing solution of 15 ml, two cleaning solution of 125 ml, 100ml deprotenizer and one quality control of 10 ml (5ml each)
- 14. Should have an alpha numeric display
- 15. Should have a minimum memory capacity of at least 100 samples
- 16. Should work on 200-240Vac 50Hz power supply
- 17. Should be supplied with off line pure sine wave UPS of sufficient capacity for a minimum back up of 30 minutes
- 18. Should be provided with calibration certificate issued by the manufacturer at the time of installation and calibration certificate should be issued for the machine by the supplier during preventive maintenance visit in the warranty/AMC period /CMC period
- 19. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission/FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid
- 20. The amount and rate of various reagents required to perform 1000test /month shall be quoted separately and this rate will be taken for evaluation
- 21. The available pack size and amount /pack (in ml) of these reagents available for each bidder shall be quoted. This rate shall be fixed for 3years
- 22. All electrodes, tubing set and membrane assembly will be covered in warranty and CAMC. The rate of each type of electrode shall be quoted separately and this rate will be fixed for 3 years if required to be procured in case of idling machines.
- 23. Facility for reagent inventory display should be available
- 24. The bidder shall provide the quantity of reagents required to perform 1000tests per month and the available standard pack size in the offer form as per format enclosed in the tender document. The price bid will not be considered for evaluation if any discrepancy in the quantity mentioned in the offer form and price bid. The quantity offered for 1000tests will be checked either in GMC Thiruvananthapuram OR State Public Health Lab, Thiruvananthapuram during demonstration. The bidder shall provide the machine and full pack of reagents for demonstration for a period of maximum one month. If any discrepancy noticed during demonstration, then the offer will not be considered and the firm will be blacklisted for a period of 3 years for misrepresentation of facts.