



Running Contract Details	
Equipment Name	Vaccum extractor
Running Contract Valid Till	26-11-2027
Tender Ref No	KMSCL/EP/T563/327A/2025
Tendered Quantity	300
Supplier Name	M/s MEDISIL ENGINEERS
GST No	33BALPS0809M1ZE
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
PLOTNO 35 DOORNO2/17 2ND STREET GAJALAKSHMI NAGAR SRINIVASAPURAM IYYAPANTHANGAL-600056	Contact Person	SEGU MOHAMED ISMAIL
	Phone	044-79616158
	Mobile No	9380520542
	Email	medisilmak@yahoo.co.in

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Vaccum extractor <i>Model & Make : MVDSO2 / YES</i>	102900 Incl.GST :5%	8094.8	110994.8			
		102900	8094.8	110994.8			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	3 rd Year	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year
Vaccum extractor							
Labour	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00	4,000.00
Comprehensive	7,000.00	7,000.00	7,000.00	7,000.00	8,000.00	8,000.00	8,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Vaccum extractor

Technical Specification

Equipment: Vaccum Extractor

1. Should have Vacuum generator, antibacterial plastic cups, silastic cup and metal cup of different sizes with necessary electrical connection cable
2. Should have cordless Foot Switch operation
3. Should generate a steady vacuum of at least 600mmHg
4. Should work on 200-240Vac 50Hz power supply
5. Should have digital display
6. Should have traction force indicating display in kilograms