

Running Contract Details	
Equipment Name	OT Table (Ophthalmic)with Chair
Running Contract Valid Till	15-05-2025
Tender Ref No	KMSCL/EP/T416(R)/88/2021
Tendered Quantity	50
Supplier Name	M/s Appasamy Associates PVT. LTD
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	OT Table (Ophthalmic) Model & Make : AAOT	165200 Incl.GST :18%	11564	176764
2	Surgeon's Chair Motorized Model & Make : AAMST-2000	72800 Incl.GST :12%	5369	78169
		238000	16933	254933

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
OT Table (Ophthalmic)							
Labour	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00
Comprehensive	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00	7,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :OT Table (Ophthalmic)

ItemName: OT TABLE (OPHTHALMIC) WITH CHAIR

A. OPHTHALMIC OPERATION TABLE MOTORISED

1. Should be motorized table with ophthalmic head piece and adjustable wrist support.
2. Should have operated on foot-pedal control for up and down movements.
3. Should provide with antibacterial, dis-infectable mattress.
4. Locks to stabilize when in use with high quality rust-proof castors.
5. Should have table top movements essential for ophthalmic surgery.
6. The table should have trendelenburg and reverse trendelenburg movement
7. Should be made of SS-304 grade

A. SURGEON'S CHAIR(Motorised)

1. Should have motorized up and down movement
2. Should have foot operated control.
3. Should have an adjustable back support.
4. Should have height adjustable / detachable arm rests.
5. Should be made up of SS304 Grade completely with upholster seat, back rest and arm rest
6. Minimum height should be 20" and maximum should be 28".
7. Should have full swivelling castors for easy mobility. Castors with brake.