



Running Contract Details	
Equipment Name	Patient Trolley with SS Stretcher
Running Contract Valid Till	30-06-2026
Tender Ref No	KMSCL/EP/T503/1468/2023
Tendered Quantity	306
Supplier Name	M/s Parkins Enterprises
GST No	32ACEPM5252G1ZW
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Police Station Road Pathanamthitta -689645	Contact Person	Ashok Mehta
	Phone	
	Mobile No	9387 222 300
	Email	parkinsenterprises@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Patient Trolley with SS Stretcher</b> <i>Model &amp; Make : MERCURY MEDICALS MAKERS</i>	20709 Incl.GST :18%	1449.63	22158.63
		<b>20709</b>	<b>1449.63</b>	<b>22158.63</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Patient Trolley with SS Stretcher**

1. Size : 2000mm (L) x 660mm (W) x 830mm (H)
2. Main frame : 32mm OD x 1.6 mm Thick M.S E.R.W
3. Supporting frame : 25mm OD x 1.6 mm Thick M.S E.R.W
4. Castor wheels with 200 mm dia.
5. Diagonal Dual lock castors
6. Stretcher top frame 25mm OD x 1.6 mm Thick M.S E.R.W
7. Sheet should be made of CRCA of 1.2mm thickness
8. Should Supply IV pole.
9. Max. Load of 1000 Kg
10. Provision for carrying oxygen cylinder.
11. Finish: powder coated with anti-microbial epoxy polyester powder.
12. Should supply with Stainless Steel stretcher and removable mattress (washable)
13. Should comply with National / International standards

#### **Note:**

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail