

Running Contract Details	
Equipment Name	Fluid Warmer (Blood Warmer)
Running Contract Valid Till	16-10-2027
Tender Ref No	KMSCL/EP/T564(R)/158A/2024
Tendered Quantity	25
Supplier Name	M/s Biocare Diagnostics
GST No	32AAOFB5438P1ZZ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
787a-1 First Floor Kamala Building Chalappuram (Po) Calicut 673002	Contact Person	Linesh.T
	Phone	
	Mobile No	9847009640
	Email	mail@biocarecalicut.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Fluid Warmer (Blood Warmer) Model & Make : QW618 /KEEWELL MEDICAL TECHNOLOGY CO. LTD.	131250 Incl.GST :5%	10325	141575
		131250	10325	141575

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Fluid Warmer (Blood Warmer)							
Labour	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00
Comprehensive	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Fluid Warmer (Blood Warmer)

Technical Specification

Equipment: Fluid Warmer (Blood Warmer)

1. Delivers blood and fluid to the patient at normothermic temperature flow rate of 50-5000ml/hr.
2. Keep the blood and fluid warm between 37-42 °C
3. Providing constant pressure for rapid infusion
4. Should be easily transportable
5. Should be able to attach to I V pole and standard electrical sockets
6. Should use dry heat technology/ multichannel counter current heat exchanger/ equivalent technology
7. Should have audible and visual alarms for Temperature
8. Should have automatic cutoff for set temperature
9. Should be easy to use and to clean
10. Calibration certificate should be issued during the installation
11. 5 disposable adult and 1 no. of pediatric warming sets should be supplied along with each machine
12. Warm up time should be less than 60 seconds
13. If available- Consumables with built in filter should be provide.
14. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.