

Running Contract Details	
Equipment Name	Serological waterbath (Digital)
Running Contract Valid Till	24-10-2026
Tender Ref No	KMSCL/EP/T539/709B/2024
Tendered Quantity	50
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Serological waterbath (Digital) <i>Model & Make : SWB-4/ LABLINE INSTRUMENTS</i>	24638.4 Incl.GST :18%	1724.69	26363.09
		24638.4	1724.69	26363.09
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Serological waterbath (Digital)				
Labour	2,000.00	2,000.00	2,000.00	
Comprehensive	5,000.00	5,000.00	6,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Serological waterbath (Digital)

Technical Specification

Equipment: Serological Water Bath (Digital)

1. Size of the bath should be 455 mm L; 300 mm W; 175mmD (Approximate) with a minimum volume of 25 Ltrs and suitable for minimum of 6 racks.
2. Double walled inside, made of Stainless Steel and outside of Mild Steel Sheet duly painted with epoxy powder coating.
3. Should have 25 mm or better thickness of mineral wool insulation on five sides.
4. Heating element made of Copper Nickel plated tabular type and approximately extend to full length of the water bath.
5. Temperature range above room temperature to 100° C with an accuracy of $\pm 0.1^{\circ}\text{C}$.
6. Temperature control by Electronic Digital Temperature Controller-cum-Indicator
7. Control panel should consist of two neon pilot lights which indicate control cycles as well as the main heater switch and temperature control knob.
8. The water bath should have perforated and removable diffuser shelves.
9. Should have Stainless Steel cover, brass drain plug, cord with plug to work on 220/230 volts A.C.
10. Must be provided with Stirrer of 1/20 H.P. Supported by High Speed motor with stainless steel stirring rod and blade provided with speed regulator.
11. Lid of water bath should be made of Stainless steel of 304 quality.
12. Traceable calibration certificate of the Instrument to be provided